

## UL TEST REPORT AND PROCEDURE

<b>Standard:</b>	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
<b>Certification Type:</b>	Component Recognition
<b>CCN:</b>	QQHM2, QQHM8 (Power Supplies, Medical and Dental)
<b>Product:</b>	Open Frame Switching Power Supply
<b>Model:</b>	CFM05SX-M, where X may be 033 - 240 or 130-1.
<b>Rating:</b>	I/P: 100-240 Vac, 47-63 Hz, 0.2-0.1 A. O/P: See Miscellaneous for details.
<b>Applicant Name and Address:</b>	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN 506 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: Selina Hsieh

Reviewed by: Conga Chen

## **COVER PAGE FOR TEST REPORT**

Product Category:	Power Supplies for Information Technology Equipment Including Electrical Business Equipment
Product Category CCN:	QQGQ2, QQGQ8
Test Procedure:	Component Recognition
Product:	Switching Power Supply
Model/Type Reference:	CFM05SXXX-Y (XXX = 033-240 for output voltage in 0.1 V; Y = any alphanumeric character or blank for marketing purposes)
Rating(s):	Input: 100-240 Vac, 0.25 A, 47-63 Hz Output: 3.3-24 Vdc, 1.25-0.22 A, see Enclosure 7-03 for details
Standards:	UL 60950-1:2003, First Edition CSA C22.2 No. 60950-1-03 1st Ed. April 1, 2003
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN 506 TAIWAN
This Report includes the following parts, in addition to this cover page:	
<ol style="list-style-type: none"><li>1. Specific Inspection Criteria</li><li>2. Specific Technical Criteria</li><li>3. Clause Verdicts</li><li>4. National Differences</li></ol>	