## **UL TEST REPORT AND PROCEDURE**

Standard:	ANSI/AAMI ES60601-1:2005 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance); CSA C22.2 No. 60601-1:08 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance)
Certification Type:	Medical Electrical Equipment
CCN:	QQHM2, QQHM8
Product:	Switching Power Supply
Model:	CFM150MX (X can be 120, 240, 360 or 480)
Rating:	I/P: 100-240Vac, 47-63Hz, 2-0.8A O/P: Model Name Output Rating CFM150M120 12 Vdc, 12.5 A CFM150M240 24 Vdc, 6.25 A CFM150M360 36 Vdc, 4.17 A CFM150M480 48 Vdc. 3.13 A
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN, 506 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

\*Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by:	Emma Xing/ Stone Zhang Underwriters Laboratories Inc.
Reviewed bv:	Calvin Tang