

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES60601-1:2005 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance); CSA C22.2 No. 60601-1:08 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance)										
Certification Type:	Medical Electrical Equipment										
CCN:	QQHM2, QQHM8										
Product:	Switching Power Supply										
Model:	CFM150MX (X can be 120, 240, 360 or 480)										
Rating:	I/P: 100-240Vac, 47-63Hz, 2-0.8A O/P: <table><thead><tr><th>Model Name</th><th>Output Rating</th></tr></thead><tbody><tr><td>CFM150M120</td><td>12 Vdc, 12.5 A</td></tr><tr><td>CFM150M240</td><td>24 Vdc, 6.25 A</td></tr><tr><td>CFM150M360</td><td>36 Vdc, 4.17 A</td></tr><tr><td>CFM150M480</td><td>48 Vdc, 3.13 A</td></tr></tbody></table>	Model Name	Output Rating	CFM150M120	12 Vdc, 12.5 A	CFM150M240	24 Vdc, 6.25 A	CFM150M360	36 Vdc, 4.17 A	CFM150M480	48 Vdc, 3.13 A
Model Name	Output Rating										
CFM150M120	12 Vdc, 12.5 A										
CFM150M240	24 Vdc, 6.25 A										
CFM150M360	36 Vdc, 4.17 A										
CFM150M480	48 Vdc, 3.13 A										
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN, 506 TAIWAN										

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

***Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.**

Prepared by: Emma Xing/ Stone Zhang
Underwriters Laboratories Inc.

Reviewed by: Calvin Tang
Underwriters Laboratories Inc.