

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES60601-1:2005 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance); CSA C22.2 No. 60601-1:08 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance)																				
Certification Type:	Medical Electrical Equipment																				
CCN:	QQHM2, QQHM8																				
Product:	Open Frame Power Supply																				
Model:	CFM40MXXX (where XXX may be 033, 050, 090, 120, 150, 240, 300, 360, 480, denoting output rating)																				
Rating:	Input: 100-240Vac, 47-63Hz, 1-0.55A Output: <table><thead><tr><th>Model Name</th><th>Output Rating</th></tr></thead><tbody><tr><td>CFM40M033</td><td>3.3 Vdc, 6 A</td></tr><tr><td>CFM40M050</td><td>5 Vdc, 6 A</td></tr><tr><td>CFM40M090</td><td>9 Vdc, 4.45 A</td></tr><tr><td>CFM40M120</td><td>12 Vdc, 3.34 A</td></tr><tr><td>CFM40M150</td><td>15 Vdc, 2.67 A</td></tr><tr><td>CFM40M240</td><td>24 Vdc, 1.67 A</td></tr><tr><td>CFM40M300</td><td>30 Vdc, 1.33 A</td></tr><tr><td>CFM40M360</td><td>36 Vdc, 1.11 A</td></tr><tr><td>CFM40M480</td><td>48 Vdc, 0.834 A</td></tr></tbody></table>	Model Name	Output Rating	CFM40M033	3.3 Vdc, 6 A	CFM40M050	5 Vdc, 6 A	CFM40M090	9 Vdc, 4.45 A	CFM40M120	12 Vdc, 3.34 A	CFM40M150	15 Vdc, 2.67 A	CFM40M240	24 Vdc, 1.67 A	CFM40M300	30 Vdc, 1.33 A	CFM40M360	36 Vdc, 1.11 A	CFM40M480	48 Vdc, 0.834 A
Model Name	Output Rating																				
CFM40M033	3.3 Vdc, 6 A																				
CFM40M050	5 Vdc, 6 A																				
CFM40M090	9 Vdc, 4.45 A																				
CFM40M120	12 Vdc, 3.34 A																				
CFM40M150	15 Vdc, 2.67 A																				
CFM40M240	24 Vdc, 1.67 A																				
CFM40M300	30 Vdc, 1.33 A																				
CFM40M360	36 Vdc, 1.11 A																				
CFM40M480	48 Vdc, 0.834 A																				
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN, 506 TAIWAN																				

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.