

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES60601-1:2005 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance); CSA C22.2 No. 60601-1:08 (Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance)
Certification Type:	Medical Electrical Equipment
CCN:	QQHM2, QQHM8
Product:	Switching Power Supply
Model:	CFM60MXXX (where XXX may be 050, 120, 150, 240, 480 for denoting output rating)
Rating:	Input: 100-240Vac, 47-63Hz, 1.6-0.8A; Output: for model CFM60M050 5Vdc, 8A for model CFM60M120 12Vdc, 5A for model CFM60M150 15Vdc, 4A for model CFM60M240 24Vdc, 2.5A for model CFM60M480 48Vdc, 1.25A
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN, 506 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Prepared by: Berman Pang/ Zang Lee
Underwriters Laboratories Inc.

Reviewed by: David Shih
Underwriters Laboratories Inc.