

## UL TEST REPORT AND PROCEDURE

<b>Standard:</b>	ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)(Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance)
<b>Certification Type:</b>	Component Recognition
<b>CCN:</b>	QQHM2, QQHM8 (Power Supplies, Medical and Dental)
<b>Product:</b>	Switching Power Adapter
<b>Model:</b>	TR60MX, where X=05, 12, 15, 18, 19, 24, 36 or 48 to represent different output ratings.
<b>Rating:</b>	I/P: 100-240Vac, 47-63Hz, 0.75-0.4A for Model TR60M05. 100-240Vac, 47-63Hz, 1.5-0.7A for Models TR60M12, TR60M15, TR60M18, TR60M19, TR60M24, TR60M36 and TR60M48.  O/P: TR60M05: 5Vdc, 6A TR60M12: 12Vdc, 5A TR60M15: 15Vdc, 4A TR60M18: 18Vdc, 3.33A TR60M19: 19Vdc, 3.15A TR60M24: 24Vdc, 2.5A TR60M36: 36Vdc, 1.66A TR60M48: 48Vdc, 1.25A
<b>Applicant Name and Address:</b>	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN 506 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.