

Certificate



Certificate no.

TA 50602985 01

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd.,
Fu Hsing Park Fu Hsing Hsiang,
Chang Hua Hsien 506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1-KM CN23T925 001

Client Reference: 48222370/ SPC

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation : TR70MAy
y stands for : 120,150,180,190,240,360 and 480
Rated Voltage : AC 100-240V, 47-63Hz
Rated Current : 1.5A
Output Voltage : see constructional dataform
Output Current : see constructional dataform
max.Ambient Temperature: 40°C
Protection Class : I

Remark:

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)

27/09/2023

Certificate



Certificate no.

TA 50602881 01

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd.,
Fu Hsing Park Fu Hsing Hsiang,
Chang Hua Hsien 506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1-KM CN238U8P 004

Client Reference: P01112699/SPC

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2
IEC 60601-1-11:2015+A1
EN 60601-1-11:2015+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation : TR70MBy
y stands for : 120,150,180,190,240,360 and 480
Rated Voltage : AC 100-240V, 47-63Hz
Rated Current : 1.5A
Output Voltage : see constructional dataform
Output Current : see constructional dataform
max.Ambient Temperature: 40°C
Protection Class : II

Remark: 1. The usability requirements of the standard were not addressed.

2. This certificate is valid for maximum 5 years.

This license sheet replaces certificate TA 50602881, sheet 01 dated 27.09.2023.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker

Date of Issue
(day/mo/yr)

18/03/2024

