

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES 60601-1: 2005, 1st Edition (Medical Electrical Equipment, Part 1: General requirements for basic safety and essential performance, includes National Differences for USA)																					
	CAN/CSA-C22.2 No. 60601-1:08, 2nd Edition (for Medical Electrical Equipment, Part 1: General requirements for basic safety and essential performance, includes National Differences for Canada)																					
Certification Type:	Component Recognition																					
CCN:	QQHM2, QQHM8																					
Product:	Adapter Power Supply																					
Model:	TRG21AXXX (where XXX may be 050, 090, 120, 150, 180 and 240)																					
Rating:	Input: 100-240Vac, 47-63Hz, 0.5-0.25A Output Rating:																					
	<table border="1"> <thead> <tr> <th>Model Name</th> <th>Vdc</th> <th>A</th> </tr> </thead> <tbody> <tr> <td>TRG21A050</td> <td>5</td> <td>3.0</td> </tr> <tr> <td>TRG21A090</td> <td>9</td> <td>2.3</td> </tr> <tr> <td>TRG21A120</td> <td>12</td> <td>1.8</td> </tr> <tr> <td>TRG21A150</td> <td>15</td> <td>1.4</td> </tr> <tr> <td>TRG21A180</td> <td>18</td> <td>1.2</td> </tr> <tr> <td>TRG21A240</td> <td>24</td> <td>0.9</td> </tr> </tbody> </table>	Model Name	Vdc	A	TRG21A050	5	3.0	TRG21A090	9	2.3	TRG21A120	12	1.8	TRG21A150	15	1.4	TRG21A180	18	1.2	TRG21A240	24	0.9
Model Name	Vdc	A																				
TRG21A050	5	3.0																				
TRG21A090	9	2.3																				
TRG21A120	12	1.8																				
TRG21A150	15	1.4																				
TRG21A180	18	1.2																				
TRG21A240	24	0.9																				
Applicant Name and Address:	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN, 506 TAIWAN																					

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: **Rios Lin / Zang Lee**
UL LLC

Reviewed by: UL LLC